

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032297.D
 Acq On : 21 Oct 2022 18:29
 Operator : JC/MD
 Sample : N5248-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-22R-20221018

Quant Time: Oct 22 05:13:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	141265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218082	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89933	48.937	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.880%
35) Dibromofluoromethane	5.391	113	68681	45.257	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	8.647	98	253551	46.415	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.820%
62) 4-Bromofluorobenzene	11.079	95	89016	44.500	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.000%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1279	2.145	ug/l #	84
44) Trichloroethene	7.129	130	192775	123.730	ug/l	95
52) Toluene	8.720	92	2441	0.666	ug/l	92
64) Tetrachloroethene	9.275	164	6206	5.013	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032297.D
Acq On : 21 Oct 2022 18:29
Operator : JC/MD
Sample : N5248-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-22R-20221018

Quant Time: Oct 22 05:13:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

