

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032298.D
 Acq On : 21 Oct 2022 18:52
 Operator : JC/MD
 Sample : N5248-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-6-20221018

Quant Time: Oct 22 05:14:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	144716	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220420	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90094	47.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.720%
35) Dibromofluoromethane	5.385	113	69011	44.993	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.980%
50) Toluene-d8	8.653	98	258043	46.736	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.480%
62) 4-Bromofluorobenzene	11.079	95	89429	44.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	833	1.363	ug/l #	80
32) 1,1,1-Trichloroethane	5.391	97	1350	0.567	ug/l #	50
44) Trichloroethene	7.129	130	47406	30.104	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032298.D
Acq On : 21 Oct 2022 18:52
Operator : JC/MD
Sample : N5248-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-6-20221018

Quant Time: Oct 22 05:14:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

