

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032302.D
 Acq On : 21 Oct 2022 20:25
 Operator : JC/MD
 Sample : N5248-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-17-20221018

Quant Time: Oct 22 05:15:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217350	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87406	47.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.440%
35) Dibromofluoromethane	5.391	113	67743	44.790	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.580%
50) Toluene-d8	8.653	98	257484	47.293	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.580%
62) 4-Bromofluorobenzene	11.079	95	88811	44.547	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	911	1.517	ug/l	90
27) cis-1,2-Dichloroethene	4.501	96	1480	0.914	ug/l	96
32) 1,1,1-Trichloroethane	5.385	97	18840	8.043	ug/l	99
44) Trichloroethene	7.135	130	10726	6.907	ug/l	99
52) Toluene	8.720	92	3599	0.985	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032302.D
Acq On : 21 Oct 2022 20:25
Operator : JC/MD
Sample : N5248-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-17-20221018

Quant Time: Oct 22 05:15:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

