

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032308.D
 Acq On : 21 Oct 2022 23:40
 Operator : JC/MD
 Sample : VX1021WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBL02

Quant Time: Oct 22 05:18:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	162207	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250210	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220903	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	108624	51.477	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.960%	
35) Dibromofluoromethane	5.385	113	83963	48.223	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.440%	
50) Toluene-d8	8.647	98	312424	49.848	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.700%	
62) 4-Bromofluorobenzene	11.079	95	113760	49.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.140%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032308.D
Acq On : 21 Oct 2022 23:40
Operator : JC/MD
Sample : VX1021WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1021WBL02

Quant Time: Oct 22 05:18:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

