

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102122\  
 Data File : VX032314.D  
 Acq On : 22 Oct 2022 02:29  
 Operator : JC/MD  
 Sample : N5248-13  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 FSND-MW-12R-20221018

Quant Time: Oct 22 06:04:54 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X101922W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 22 04:56:40 2022  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc    | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|---------|--------|----------|
| -----                        |                |      |            |         |        |          |
| Internal Standards           |                |      |            |         |        |          |
| 1) Pentafluorobenzene        | 5.556          | 168  | 146884     | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763          | 114  | 230307     | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055         | 117  | 200010     | 50.000  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024         | 152  | 89195      | 50.000  | ug/l   | 0.00     |
|                              |                |      |            |         |        |          |
| System Monitoring Compounds  |                |      |            |         |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.958          | 65   | 92048      | 48.172  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 78 - 117 |      | Recovery = | 96.340% |        |          |
| 35) Dibromofluoromethane     | 5.385          | 113  | 72409      | 45.181  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 75 - 124 |      | Recovery = | 90.360% |        |          |
| 50) Toluene-d8               | 8.647          | 98   | 268107     | 46.474  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 92 - 112 |      | Recovery = | 92.940% |        |          |
| 62) 4-Bromofluorobenzene     | 11.079         | 95   | 93711      | 44.360  | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 83 - 123 |      | Recovery = | 88.720% |        |          |
|                              |                |      |            |         |        |          |
| Target Compounds             |                |      |            |         |        |          |
| 16) Acetone                  | 2.392          | 43   | 1450       | 2.338   | ug/l   | 93       |
| 21) trans-1,2-Dichloroethene | 3.099          | 96   | 372        | 0.261   | ug/l # | 70       |
| 30) Chloroform               | 5.098          | 83   | 905        | 0.319   | ug/l # | 63       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 5177       | 2.141   | ug/l   | 95       |
| 38) Carbon Tetrachloride     | 5.684          | 117  | 7783       | 3.576   | ug/l   | 97       |
| 44) Trichloroethene          | 7.129          | 130  | 11239      | 6.831   | ug/l   | 98       |
| 52) Toluene                  | 8.720          | 92   | 2663       | 0.688   | ug/l   | 98       |
| 64) Tetrachloroethene        | 9.275          | 164  | 1718       | 1.301   | ug/l   | 96       |
| -----                        |                |      |            |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102122\  
Data File : VX032314.D  
Acq On : 22 Oct 2022 02:29  
Operator : JC/MD  
Sample : N5248-13  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 38 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
FSND-MW-12R-20221018

Quant Time: Oct 22 06:04:54 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X101922W.M  
Quant Title : SW846 8260  
QLast Update : Sat Oct 22 04:56:40 2022  
Response via : Initial Calibration

