

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032318.D
 Acq On : 22 Oct 2022 04:02
 Operator : JC/MD
 Sample : N5248-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-34-20221019

Quant Time: Oct 22 06:05:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227717	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195275	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90822	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91766	48.444	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.880%		
35) Dibromofluoromethane	5.385	113	71113	44.877	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.760%		
50) Toluene-d8	8.653	98	264714	46.408	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	92.820%		
62) 4-Bromofluorobenzene	11.079	95	91574	43.841	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	87.680%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	1567	0.945	ug/l	80
44) Trichloroethene	7.128	130	36643	22.524	ug/l	100
52) Toluene	8.720	92	2742	0.716	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032318.D
Acq On : 22 Oct 2022 04:02
Operator : JC/MD
Sample : N5248-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-34-20221019

Quant Time: Oct 22 06:05:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

