

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102122\
 Data File : VX032326.D
 Acq On : 22 Oct 2022 07:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 07:31:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.457	0.461	-0.9	102	0.00
3 P	Chloromethane	0.482	0.407	15.6	86	0.00
4 C	Vinyl Chloride	0.493	0.466	5.5#	90	0.00
5 T	Bromomethane	0.516	0.334	35.3#	62	0.00
6 T	Chloroethane	0.395	0.381	3.5	88	0.00
7 T	Trichlorofluoromethane	0.876	0.937	-7.0	107	0.00
8 T	Diethyl Ether	0.298	0.304	-2.0	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.475	0.485	-2.1	104	0.00
10 T	Methyl Iodide	0.686	0.502	26.8#	68	0.00
11 T	Tert butyl alcohol	0.105	0.076	27.6#	81	0.04
12 CM	1,1-Dichloroethene	0.446	0.452	-1.3#	97	0.00
13 T	Acrolein	0.079	0.082	-3.8	103	0.00
14 T	Allyl chloride	0.701	0.742	-5.8	100	0.00
15 T	Acrylonitrile	0.244	0.236	3.3	92	0.00
16 T	Acetone	0.211	0.211	0.0	99	0.01
17 T	Carbon Disulfide	1.045	0.975	6.7	88	0.00
18 T	Methyl Acetate	0.717	0.716	0.1	95	0.00
19 T	Methyl tert-butyl Ether	1.538	1.683	-9.4	105	0.00
20 T	Methylene Chloride	0.542	0.518	4.4	97	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.489	-0.6	95	0.00
22 T	Diisopropyl ether	1.535	1.672	-8.9	105	0.00
23 T	Vinyl Acetate	1.227	1.321	-7.7	101	0.00
24 P	1,1-Dichloroethane	0.887	0.938	-5.7	103	0.00
25 T	2-Butanone	0.330	0.322	2.4	94	0.00
26 T	2,2-Dichloropropane	0.639	0.596	6.7	86	0.00
27 T	cis-1,2-Dichloroethene	0.569	0.599	-5.3	99	0.00
28 T	Bromochloromethane	0.366	0.354	3.3	100	0.00
29 T	Tetrahydrofuran	0.213	0.203	4.7	90	0.00
30 C	Chloroform	0.965	1.047	-8.5#	104	0.00
31 T	Cyclohexane	0.744	0.758	-1.9	100	0.00
32 T	1,1,1-Trichloroethane	0.823	0.954	-15.9	108	0.00
33 S	1,2-Dichloroethane-d4	0.650	0.692	-6.5	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.348	0.348	0.0	98	0.00
36 T	1,1-Dichloropropene	0.448	0.452	-0.9	101	0.00
37 T	Ethyl Acetate	0.438	0.418	4.6	96	0.00
38 T	Carbon Tetrachloride	0.472	0.518	-9.7	106	0.00
39 T	Methylcyclohexane	0.535	0.522	2.4	99	0.00
40 TM	Benzene	1.297	1.284	1.0	98	0.00
41 T	Methacrylonitrile	0.231	0.234	-1.3	94	0.00
42 TM	1,2-Dichloroethane	0.497	0.545	-9.7	108	0.00
43 T	Isopropyl Acetate	0.688	0.720	-4.7	100	0.00
44 TM	Trichloroethene	0.357	0.355	0.6	98	0.00
45 C	1,2-Dichloropropane	0.327	0.336	-2.8#	100	0.00
46 T	Dibromomethane	0.244	0.251	-2.9	100	0.00
47 T	Bromodichloromethane	0.451	0.497	-10.2	106	0.00
48 T	Methyl methacrylate	0.343	0.361	-5.2	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
 Data File : VX032326.D
 Acq On : 22 Oct 2022 07:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 07:31:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.006	14.3	88	0.02
50 S	Toluene-d8	1.252	1.235	1.4	94	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.431	0.0	95	0.00
52 CM	Toluene	0.840	0.848	-1.0#	99	0.00
53 T	t-1,3-Dichloropropene	0.451	0.504	-11.8	103	0.00
54 T	cis-1,3-Dichloropropene	0.502	0.534	-6.4	100	0.00
55 T	1,1,2-Trichloroethane	0.341	0.357	-4.7	101	0.00
56 T	Ethyl methacrylate	0.492	0.528	-7.3	102	0.00
57 T	1,3-Dichloropropane	0.553	0.576	-4.2	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.225	0.212	5.8	92	0.00
59 T	2-Hexanone	0.322	0.321	0.3	95	0.00
60 T	Dibromochloromethane	0.350	0.387	-10.6	103	0.00
61 T	1,2-Dibromoethane	0.362	0.377	-4.1	100	0.00
62 S	4-Bromofluorobenzene	0.459	0.499	-8.7	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.330	0.317	3.9	98	0.00
65 PM	Chlorobenzene	1.025	1.018	0.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.367	0.395	-7.6	104	0.00
67 C	Ethyl Benzene	1.815	1.858	-2.4#	102	0.00
68 T	m/p-Xylenes	0.717	0.716	0.1	100	0.00
69 T	o-Xylene	0.687	0.708	-3.1	102	0.00
70 T	Styrene	1.158	1.204	-4.0	103	0.00
71 P	Bromoform	0.278	0.298	-7.2	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	3.568	3.599	-0.9	103	0.00
74 T	N-amyl acetate	1.318	1.382	-4.9	101	0.00
75 P	1,1,2,2-Tetrachloroethane	1.147	1.119	2.4	100	0.00
76 T	1,2,3-Trichloropropane	0.991	0.967	2.4	100	0.00
77 T	Bromobenzene	0.875	0.879	-0.5	103	0.00
78 T	n-propylbenzene	4.085	4.205	-2.9	103	0.00
79 T	2-Chlorotoluene	2.462	2.540	-3.2	106	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.173	-3.0	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.280	3.8	93	0.00
82 T	4-Chlorotoluene	2.867	3.000	-4.6	106	0.00
83 T	tert-Butylbenzene	3.083	3.174	-3.0	105	0.00
84 T	1,2,4-Trimethylbenzene	2.996	3.157	-5.4	105	0.00
85 T	sec-Butylbenzene	3.726	3.848	-3.3	104	0.00
86 T	p-Isopropyltoluene	3.171	3.322	-4.8	105	0.00
87 T	1,3-Dichlorobenzene	1.683	1.690	-0.4	104	0.00
88 T	1,4-Dichlorobenzene	1.729	1.717	0.7	103	0.00
89 T	n-Butylbenzene	2.793	2.857	-2.3	101	0.00
90 T	Hexachloroethane	0.492	0.547	-11.2	110	0.00
91 T	1,2-Dichlorobenzene	1.652	1.694	-2.5	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.219	0.230	-5.0	98	0.00
93 T	1,2,4-Trichlorobenzene	1.076	1.057	1.8	100	0.00
94 T	Hexachlorobutadiene	0.467	0.433	7.3	97	0.00
95 T	Naphthalene	3.445	3.443	0.1	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102122\
Data File : VX032326.D
Acq On : 22 Oct 2022 07:06
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 07:31:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.082	1.046	3.3	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6