

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043456.D
 Acq On : 21 Oct 2024 15:49
 Operator : JC/MD
 Sample : P4430-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:40:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	100546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191497	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86119	52.445	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.900%
35) Dibromofluoromethane	5.385	113	63491	48.076	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.160%
50) Toluene-d8	8.647	98	225502	49.273	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.540%
62) 4-Bromofluorobenzene	11.079	95	84082	50.571	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	28059	40.918	ug/l	96
18) Methyl Acetate	2.715	43	3111	1.808	ug/l	100
20) Methylene Chloride	2.794	84	2805	2.219	ug/l	91
43) Isopropyl Acetate	6.342	43	95798	28.948	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043456.D
Acq On : 21 Oct 2024 15:49
Operator : JC/MD
Sample : P4430-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:40:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

