

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043461.D
 Acq On : 21 Oct 2024 17:44
 Operator : JC/MD
 Sample : PB164196ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:42:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	90505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	168248	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64818	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	76007	51.423	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.840%		
35) Dibromofluoromethane	5.385	113	55718	48.020	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.040%		
50) Toluene-d8	8.647	98	199721	49.670	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.340%		
62) 4-Bromofluorobenzene	11.079	95	72987	49.964	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 99.920%		
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	26864	43.521	ug/l	99
18) Methyl Acetate	2.715	43	3419	2.208	ug/l	95
20) Methylene Chloride	2.794	84	3436	3.019	ug/l	92
43) Isopropyl Acetate	6.349	43	99060	34.070	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043461.D
Acq On : 21 Oct 2024 17:44
Operator : JC/MD
Sample : PB164196ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:42:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

