

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043463.D
 Acq On : 21 Oct 2024 18:30
 Operator : JC/MD
 Sample : PB164196ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:43:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96300	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	180394	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	158578	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	71319	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82686	52.575	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.140%
35) Dibromofluoromethane	5.391	113	59595	47.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.800%
50) Toluene-d8	8.647	98	217570	50.465	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	11.079	95	79554	50.792	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.580%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	28252	43.016	ug/l	99
18) Methyl Acetate	2.715	43	3812	2.314	ug/l	91
20) Methylene Chloride	2.788	84	3194	2.638	ug/l	89
43) Isopropyl Acetate	6.342	43	102609	32.915	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043463.D
Acq On : 21 Oct 2024 18:30
Operator : JC/MD
Sample : PB164196ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:43:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

