

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043464.D
 Acq On : 21 Oct 2024 18:54
 Operator : JC/MD
 Sample : PB164196ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:43:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	93088	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173226	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	149117	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66719	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78636	51.725	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.460%
35) Dibromofluoromethane	5.385	113	57427	48.071	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.140%
50) Toluene-d8	8.647	98	204493	49.395	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.780%
62) 4-Bromofluorobenzene	11.079	95	73257	48.707	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.420%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	27036	42.585	ug/l	96
18) Methyl Acetate	2.715	43	3728	2.341	ug/l	94
20) Methylene Chloride	2.794	84	3700	3.161	ug/l #	86
43) Isopropyl Acetate	6.342	43	96588	32.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043464.D
Acq On : 21 Oct 2024 18:54
Operator : JC/MD
Sample : PB164196ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:43:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

