

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043465.D
 Acq On : 21 Oct 2024 19:17
 Operator : JC/MD
 Sample : PB164196ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:44:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	91407	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	170691	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	152813	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68156	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78694	52.715	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.440%
35) Dibromofluoromethane	5.391	113	57980	49.254	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.500%
50) Toluene-d8	8.647	98	205923	50.479	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	76584	51.676	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.360%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	27340	43.855	ug/l	99
18) Methyl Acetate	2.709	43	3664	2.343	ug/l	98
20) Methylene Chloride	2.794	84	3386	2.946	ug/l	96
43) Isopropyl Acetate	6.342	43	103066	34.941	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043465.D
Acq On : 21 Oct 2024 19:17
Operator : JC/MD
Sample : PB164196ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:44:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

