

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043466.D
 Acq On : 21 Oct 2024 19:40
 Operator : JC/MD
 Sample : P4443-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 03:54:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	94463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	174832	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	155038	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76955	49.883	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.760%
35) Dibromofluoromethane	5.391	113	56428	46.801	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.600%
50) Toluene-d8	8.647	98	210337	50.340	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	11.079	95	73975	48.733	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	37119	57.615	ug/l	94
18) Methyl Acetate	2.715	43	4003	2.477	ug/l	92
20) Methylene Chloride	2.794	84	2206	1.857	ug/l #	89
43) Isopropyl Acetate	6.348	43	43093	14.263	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043466.D
Acq On : 21 Oct 2024 19:40
Operator : JC/MD
Sample : P4443-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 03:54:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

