

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043467.D
 Acq On : 21 Oct 2024 20:03
 Operator : JC/MD
 Sample : P4443-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:45:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	176510	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	157984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66031	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79895	51.813	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	103.620%	
35) Dibromofluoromethane	5.385	113	59286	48.704	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.400%	
50) Toluene-d8	8.647	98	210204	49.830	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.660%	
62) 4-Bromofluorobenzene	11.079	95	79197	51.677	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	103.360%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	32138	49.908	ug/l	96
18) Methyl Acetate	2.709	43	3379	2.092	ug/l #	90
20) Methylene Chloride	2.794	84	3077	2.592	ug/l #	88
43) Isopropyl Acetate	6.348	43	61019	20.004	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043467.D
Acq On : 21 Oct 2024 20:03
Operator : JC/MD
Sample : P4443-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:45:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

