

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043468.D
 Acq On : 21 Oct 2024 20:26
 Operator : JC/MD
 Sample : P4458-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 22 02:45:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96342	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	181205	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	162177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72834	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86296	54.846	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.700%	
35) Dibromofluoromethane	5.391	113	62029	49.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.280%	
50) Toluene-d8	8.647	98	215764	49.823	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.640%	
62) 4-Bromofluorobenzene	11.079	95	81114	51.557	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.120%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	24810	37.759	ug/l	100
18) Methyl Acetate	2.709	43	3671	2.227	ug/l	95
20) Methylene Chloride	2.788	84	3039	2.508	ug/l #	81
43) Isopropyl Acetate	6.342	43	98823	31.559	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043468.D
Acq On : 21 Oct 2024 20:26
Operator : JC/MD
Sample : P4458-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 22 02:45:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

