

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102124\
 Data File : VX043446.D
 Acq On : 21 Oct 2024 11:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 17:40:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.582	0.543	6.7	80	0.00
3 P	Chloromethane	0.599	0.623	-4.0	92	0.00
4 C	Vinyl Chloride	0.600	0.590	1.7#	87	0.00
5 T	Bromomethane	0.245	0.285	-16.3	106	0.00
6 T	Chloroethane	0.223	0.351	-57.4#	138	0.00
7 T	Trichlorofluoromethane	0.896	0.942	-5.1	96	0.00
8 T	Diethyl Ether	0.349	0.380	-8.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.559	0.554	0.9	88	0.00
10 T	Methyl Iodide	0.703	0.769	-9.4	95	0.00
11 T	Tert butyl alcohol	0.155	0.172	-11.0	95	-0.01
12 CM	1,1-Dichloroethene	0.556	0.533	4.1#	86	0.00
13 T	Acrolein	0.147	0.140	4.8	84	0.00
14 T	Allyl chloride	0.913	1.054	-15.4	102	0.00
15 T	Acrylonitrile	0.317	0.350	-10.4	97	0.00
16 T	Acetone	0.341	0.400	-17.3	112	0.00
17 T	Carbon Disulfide	1.351	1.172	13.2	79	0.00
18 T	Methyl Acetate	0.855	0.935	-9.4	93	0.00
19 T	Methyl tert-butyl Ether	1.987	2.229	-12.2	99	0.00
20 T	Methylene Chloride	0.629	0.680	-8.1	99	0.00
21 T	trans-1,2-Dichloroethene	0.553	0.591	-6.9	95	0.00
22 T	Diisopropyl ether	1.836	2.205	-20.1	106	0.00
23 T	Vinyl Acetate	1.940	1.799	7.3	71	0.00
24 P	1,1-Dichloroethane	1.069	1.237	-15.7	102	0.00
25 T	2-Butanone	0.468	0.534	-14.1	101	0.00
26 T	2,2-Dichloropropane	1.018	1.137	-11.7	100	0.00
27 T	cis-1,2-Dichloroethene	0.690	0.793	-14.9	103	0.00
28 T	Bromochloromethane	0.422	0.433	-2.6	95	0.00
29 T	Tetrahydrofuran	0.298	0.314	-5.4	95	0.00
30 C	Chloroform	1.174	1.388	-18.2#	107	0.00
31 T	Cyclohexane	0.881	0.847	3.9	87	0.00
32 T	1,1,1-Trichloroethane	1.079	1.131	-4.8	93	0.00
33 S	1,2-Dichloroethane-d4	0.817	0.823	-0.7	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
35 S	Dibromofluoromethane	0.345	0.339	1.7	96	0.00
36 T	1,1-Dichloropropene	0.436	0.411	5.7	91	0.00
37 T	Ethyl Acetate	0.523	0.504	3.6	93	0.00
38 T	Carbon Tetrachloride	0.516	0.482	6.6	88	0.00
39 T	Methylcyclohexane	0.548	0.499	8.9	87	0.00
40 TM	Benzene	1.321	1.364	-3.3	100	0.00
41 T	Methacrylonitrile	0.272	0.277	-1.8	92	0.00
42 TM	1,2-Dichloroethane	0.538	0.565	-5.0	100	0.00
43 T	Isopropyl Acetate	0.864	0.860	0.5	93	0.00
44 TM	Trichloroethene	0.350	0.339	3.1	95	0.00
45 C	1,2-Dichloropropane	0.324	0.352	-8.6#	103	0.00
46 T	Dibromomethane	0.260	0.264	-1.5	97	0.00
47 T	Bromodichloromethane	0.511	0.542	-6.1	99	0.00
48 T	Methyl methacrylate	0.424	0.415	2.1	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
 Data File : VX043446.D
 Acq On : 21 Oct 2024 11:43
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 21 17:40:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	84	0.00
50 S	Toluene-d8	1.195	1.099	8.0	87	0.00
51 T	4-Methyl-2-Pentanone	0.525	0.525	0.0	93	0.00
52 CM	Toluene	0.828	0.855	-3.3#	97	0.00
53 T	t-1,3-Dichloropropene	0.507	0.538	-6.1	95	0.00
54 T	cis-1,3-Dichloropropene	0.544	0.586	-7.7	101	0.00
55 T	1,1,2-Trichloroethane	0.329	0.344	-4.6	97	0.00
56 T	Ethyl methacrylate	0.538	0.557	-3.5	92	0.00
57 T	1,3-Dichloropropane	0.560	0.575	-2.7	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.239	1.2	87	0.00
59 T	2-Hexanone	0.397	0.405	-2.0	94	0.00
60 T	Dibromochloromethane	0.382	0.393	-2.9	93	0.00
61 T	1,2-Dibromoethane	0.353	0.354	-0.3	93	0.00
62 S	4-Bromofluorobenzene	0.434	0.433	0.2	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.347	0.329	5.2	93	0.00
65 PM	Chlorobenzene	1.075	1.115	-3.7	103	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.401	-5.5	100	0.00
67 C	Ethyl Benzene	1.858	1.837	1.1#	95	0.00
68 T	m/p-Xylenes	0.702	0.698	0.6	97	0.00
69 T	o-Xylene	0.698	0.714	-2.3	100	0.00
70 T	Styrene	1.131	1.213	-7.3	101	0.00
71 P	Bromoform	0.291	0.273	6.2	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.887	3.643	6.3	93	0.00
74 T	N-amyl acetate	1.845	1.613	12.6	88	0.00
75 P	1,1,2,2-Tetrachloroethane	1.266	1.162	8.2	93	0.00
76 T	1,2,3-Trichloropropane	1.109	1.017	8.3	95	0.00
77 T	Bromobenzene	0.946	0.921	2.6	99	0.00
78 T	n-propylbenzene	4.268	4.144	2.9	94	0.00
79 T	2-Chlorotoluene	2.744	2.735	0.3	103	0.00
80 T	1,3,5-Trimethylbenzene	3.207	3.144	2.0	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.428	0.364	15.0	87	0.00
82 T	4-Chlorotoluene	3.125	3.175	-1.6	103	0.00
83 T	tert-Butylbenzene	3.304	3.080	6.8	94	0.00
84 T	1,2,4-Trimethylbenzene	3.226	3.237	-0.3	100	0.00
85 T	sec-Butylbenzene	3.916	3.589	8.4	91	0.00
86 T	p-Isopropyltoluene	3.267	3.105	5.0	94	0.00
87 T	1,3-Dichlorobenzene	1.666	1.671	-0.3	103	0.00
88 T	1,4-Dichlorobenzene	1.718	1.683	2.0	103	0.00
89 T	n-Butylbenzene	2.735	2.622	4.1	93	0.00
90 T	Hexachloroethane	0.613	0.559	8.8	89	0.00
91 T	1,2-Dichlorobenzene	1.704	1.657	2.8	101	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.309	0.270	12.6	87	0.00
93 T	1,2,4-Trichlorobenzene	0.983	0.995	-1.2	100	0.00
94 T	Hexachlorobutadiene	0.389	0.382	1.8	95	0.00
95 T	Naphthalene	3.612	3.161	12.5	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102124\
Data File : VX043446.D
Acq On : 21 Oct 2024 11:43
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 21 17:40:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.014	0.970	4.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6