

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048253.D
 Acq On : 21 Oct 2025 09:12
 Operator : JC/MD
 Sample : VX1021MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021MBL01

Quant Time: Oct 24 04:22:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	91961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	178643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	183312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	90101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	62909	48.791	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.580%
35) Dibromofluoromethane	5.367	113	54167	44.414	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.820%
50) Toluene-d8	8.635	98	188946	44.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.400%#
62) 4-Bromofluorobenzene	11.061	95	80532	49.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048253.D
Acq On : 21 Oct 2025 09:12
Operator : JC/MD
Sample : VX1021MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1021MBL01

Quant Time: Oct 24 04:22:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

