

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048254.D
 Acq On : 21 Oct 2025 09:40
 Operator : JC/MD
 Sample : VX1021WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1021WBL01

Quant Time: Oct 24 04:22:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	121915	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.745	114	240535	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	247603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	126551	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	89687	52.469	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.940%
35) Dibromofluoromethane	5.367	113	77294	47.070	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.140%
50) Toluene-d8	8.629	98	267020	46.391	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.780%
62) 4-Bromofluorobenzene	11.061	95	115293	52.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048254.D
Acq On : 21 Oct 2025 09:40
Operator : JC/MD
Sample : VX1021WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1021WBL01

Quant Time: Oct 24 04:22:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

