

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048257.D
 Acq On : 21 Oct 2025 10:54
 Operator : JC/MD
 Sample : Q3371-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB184-HYD-20251010

Quant Time: Oct 24 04:25:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	90641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	171684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	176825	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	86166	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	61446	48.350	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.700%
35) Dibromofluoromethane	5.373	113	54394	46.408	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.820%
50) Toluene-d8	8.634	98	184640	44.944	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.880%#
62) 4-Bromofluorobenzene	11.061	95	79196	50.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048257.D
Acq On : 21 Oct 2025 10:54
Operator : JC/MD
Sample : Q3371-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB184-HYD-20251010

Quant Time: Oct 24 04:25:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

