

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048262.D
 Acq On : 21 Oct 2025 12:37
 Operator : JC/MD
 Sample : PB170148TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB170148TB

Quant Time: Oct 21 12:55:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	111638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	224447	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	249873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	125850	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	86412	55.207	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.420%
35) Dibromofluoromethane	5.373	113	71622	46.742	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.480%
50) Toluene-d8	8.634	98	265270	49.391	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.780%
62) 4-Bromofluorobenzene	11.061	95	115506	56.531	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.060%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	20914	38.948	ug/l	99
18) Methyl Acetate	2.702	43	3831	3.416	ug/l	99
20) Methylene Chloride	2.794	84	5546	4.239	ug/l	92
25) 2-Butanone	4.550	43	3854	5.670	ug/l #	89
43) Isopropyl Acetate	6.330	43	6852	2.311	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048262.D
Acq On : 21 Oct 2025 12:37
Operator : JC/MD
Sample : PB170148TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB170148TB

Quant Time: Oct 21 12:55:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

