

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048266.D
 Acq On : 21 Oct 2025 14:00
 Operator : JC/MD
 Sample : Q3374-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-5-WC

Quant Time: Oct 24 04:27:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	127477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	252810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	265129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96921	54.227	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.460%
35) Dibromofluoromethane	5.373	113	82084	47.560	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	8.634	98	286845	47.416	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.840%
62) 4-Bromofluorobenzene	11.061	95	125243	54.420	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.840%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	11413	18.614	ug/l	93
18) Methyl Acetate	2.703	43	4208	3.286	ug/l	96
20) Methylene Chloride	2.794	84	6079	4.069	ug/l #	84
25) 2-Butanone	4.568	43	5790	7.460	ug/l	100
43) Isopropyl Acetate	6.330	43	6589	1.973	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048266.D
Acq On : 21 Oct 2025 14:00
Operator : JC/MD
Sample : Q3374-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-5-WC

Quant Time: Oct 24 04:27:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

