

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048269.D
 Acq On : 21 Oct 2025 15:02
 Operator : JC/MD
 Sample : Q3381-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR200059

Quant Time: Oct 21 15:20:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113130	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	222761	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	240301	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	86614	54.606	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.220%		
35) Dibromofluoromethane	5.373	113	73099	48.067	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.140%		
50) Toluene-d8	8.635	98	253548	47.566	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 95.140%		
62) 4-Bromofluorobenzene	11.061	95	110865	54.671	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 109.340%		
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	16660	30.617	ug/l	95
18) Methyl Acetate	2.709	43	2587	2.276	ug/l #	89
20) Methylene Chloride	2.788	84	2981	2.248	ug/l	91
43) Isopropyl Acetate	6.330	43	7785	2.645	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048269.D
Acq On : 21 Oct 2025 15:02
Operator : JC/MD
Sample : Q3381-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR200059

Quant Time: Oct 21 15:20:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

