

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048271.D
 Acq On : 21 Oct 2025 15:44
 Operator : JC/MD
 Sample : Q3392-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-20

Quant Time: Oct 24 04:27:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114140	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	243985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	259738	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130335	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	89456	55.899	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.800%
35) Dibromofluoromethane	5.379	113	76308	45.812	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	8.635	98	280684	48.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.160%
62) 4-Bromofluorobenzene	11.061	95	121948	54.905	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.800%
Target Compounds						
					Qvalue	
16) Acetone	2.373	43	26517	48.300	ug/l	97
18) Methyl Acetate	2.703	43	2855	2.490	ug/l #	69
20) Methylene Chloride	2.794	84	2941	2.199	ug/l	87
43) Isopropyl Acetate	6.324	43	7347	2.279	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048271.D
Acq On : 21 Oct 2025 15:44
Operator : JC/MD
Sample : Q3392-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-20

Quant Time: Oct 24 04:27:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

