

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048272.D
 Acq On : 21 Oct 2025 16:04
 Operator : JC/MD
 Sample : Q3392-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-19

Quant Time: Oct 24 04:28:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	104161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	215935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	243754	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122293	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	82433	56.445	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.900%
35) Dibromofluoromethane	5.373	113	66998	45.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.900%
50) Toluene-d8	8.635	98	255253	49.399	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.800%
62) 4-Bromofluorobenzene	11.061	95	112149	57.052	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.100%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	20209	40.337	ug/l	96
18) Methyl Acetate	2.703	43	2996	2.863	ug/l	93
20) Methylene Chloride	2.794	84	3169	2.596	ug/l #	82
25) 2-Butanone	4.556	43	3944	6.219	ug/l	91
43) Isopropyl Acetate	6.336	43	6901	2.419	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048272.D
Acq On : 21 Oct 2025 16:04
Operator : JC/MD
Sample : Q3392-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-19

Quant Time: Oct 24 04:28:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

