

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
 Data File : VX048277.D
 Acq On : 21 Oct 2025 17:48
 Operator : JC/MD
 Sample : Q3376-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PR132-FB01-20251016

Quant Time: Oct 24 04:29:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	95980	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	184822	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	196926	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	105671	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	78641	58.438	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.880%
35) Dibromofluoromethane	5.373	113	62781	49.756	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.520%
50) Toluene-d8	8.635	98	212078	47.953	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.900%
62) 4-Bromofluorobenzene	11.061	95	96027	57.074	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.140%
Target Compounds						
16) Acetone	2.374	43	2619	5.673	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102125\
Data File : VX048277.D
Acq On : 21 Oct 2025 17:48
Operator : JC/MD
Sample : Q3376-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PR132-FB01-20251016

Quant Time: Oct 24 04:29:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

