

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005467.D
 Acq On : 22 Oct 2018 13:45
 Operator : JC/MD
 Sample : VX1022WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1022WBL01

Quant Time: Oct 23 01:40:01 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	273242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	388576	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.12	117	329611	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149204	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	157681	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
35) Dibromofluoromethane	5.50	113	134819	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
50) Toluene-d8	8.71	98	407339	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
62) 4-Bromofluorobenzene	11.14	95	134082	40.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005467.D
Acq On : 22 Oct 2018 13:45
Operator : JC/MD
Sample : VX1022WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1022WBL01

Quant Time: Oct 23 01:40:01 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

