

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005469.D
 Acq On : 22 Oct 2018 14:46
 Operator : JC/MD
 Sample : J5610-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Oct 23 11:41:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	249513	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	350384	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145551	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	145734	52.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.26%	
35) Dibromofluoromethane	5.49	113	116088	50.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.04%	
50) Toluene-d8	8.72	98	411760	51.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.36%	
62) 4-Bromofluorobenzene	11.14	95	132778	44.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.72%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005469.D
Acq On : 22 Oct 2018 14:46
Operator : JC/MD
Sample : J5610-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Oct 23 11:41:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

