

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005476.D
 Acq On : 22 Oct 2018 17:53
 Operator : JC/MD
 Sample : PB114041ZHE#06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114041ZHE#06

Quant Time: Oct 23 02:18:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225930	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	330998	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	306596	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143920	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	149508	59.62	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	119.24%	
35) Dibromofluoromethane	5.51	113	112414	51.28	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	8.72	98	404857	53.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.58%	
62) 4-Bromofluorobenzene	11.14	95	132225	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
16) Acetone	2.45	43	6562	4.526	ug/l	97
18) Methyl Acetate	2.78	43	8918	2.162	ug/l	98
43) Isopropyl Acetate	6.46	43	163698	29.808	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005476.D
Acq On : 22 Oct 2018 17:53
Operator : JC/MD
Sample : PB114041ZHE#06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114041ZHE#06

Quant Time: Oct 23 02:18:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

