

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005484.D
 Acq On : 22 Oct 2018 21:27
 Operator : JC/MD
 Sample : J5538-17 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NL-7

Quant Time: Oct 23 03:15:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	234656	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	341666	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	294656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	141352	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	146320	56.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.36%	
35) Dibromofluoromethane	5.51	113	121695	53.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.56%	
50) Toluene-d8	8.71	98	389918	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.14	95	128771	44.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.24%	
Target Compounds						
16) Acetone	2.45	43	7959	5.286	ug/l	89
18) Methyl Acetate	2.78	43	5338	1.246	ug/l	93
43) Isopropyl Acetate	6.46	43	27783	4.901	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005484.D
Acq On : 22 Oct 2018 21:27
Operator : JC/MD
Sample : J5538-17 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NL-7

Quant Time: Oct 23 03:15:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

