

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005485.D
 Acq On : 22 Oct 2018 21:54
 Operator : JC/MD
 Sample : J5538-23 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BS01

Quant Time: Oct 23 03:23:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	232470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	325583	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	288932	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146644	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	139005	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
35) Dibromofluoromethane	5.50	113	108781	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	8.72	98	387857	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
62) 4-Bromofluorobenzene	11.14	95	130981	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
Target Compounds						
16) Acetone	2.45	43	25306	16.964	ug/l	99
20) Methylene Chloride	2.86	84	15614	6.033	ug/l	90
25) 2-Butanone	4.71	43	3812	1.747	ug/l #	84
43) Isopropyl Acetate	6.46	43	13529	2.505	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005485.D
Acq On : 22 Oct 2018 21:54
Operator : JC/MD
Sample : J5538-23 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BS01

Quant Time: Oct 23 03:23:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

