

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005491.D
 Acq On : 23 Oct 2018 01:26
 Operator : JC/MD
 Sample : VX1022WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1022WBL02

Quant Time: Oct 23 03:43:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	218420	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	325394	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	294644	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145223	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	144151	59.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	118.94%	
35) Dibromofluoromethane	5.51	113	112954	52.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.82%	
50) Toluene-d8	8.71	98	394381	53.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.60%	
62) 4-Bromofluorobenzene	11.14	95	130003	46.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005491.D
Acq On : 23 Oct 2018 01:26
Operator : JC/MD
Sample : VX1022WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1022WBL02

Quant Time: Oct 23 03:43:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

