

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005495.D
 Acq On : 23 Oct 2018 03:13
 Operator : JC/MD
 Sample : J5626-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2018.10.19-SW

Quant Time: Oct 23 13:14:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	228158	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	339621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	286804	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138122	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	143511	56.67	ug/l	-0.01
Spiked Amount						
			Recovery	=		113.34%
35) Dibromofluoromethane	5.50	113	110746	49.23	ug/l	0.00
Spiked Amount						
			Recovery	=		98.46%
50) Toluene-d8	8.72	98	383774	49.70	ug/l	0.00
Spiked Amount						
			Recovery	=		99.40%
62) 4-Bromofluorobenzene	11.14	95	125928	43.41	ug/l	0.00
Spiked Amount						
			Recovery	=		86.82%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005495.D
Acq On : 23 Oct 2018 03:13
Operator : JC/MD
Sample : J5626-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2018.10.19-SW

Quant Time: Oct 23 13:14:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

