

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\  
 Data File : VX005500.D  
 Acq On : 23 Oct 2018 05:26  
 Operator : JC/MD  
 Sample : J5621-09  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 PA-SW-01

Quant Time: Oct 23 13:18:25 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X101218W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Oct 12 11:28:43 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.66  | 168  | 226736   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.86  | 114  | 320534   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.12 | 117  | 286214   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.08 | 152  | 134613   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| 33) 1,2-Dichloroethane-d4   | 6.07   | 65  | 140158   | 55.70 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 111.40% |      |
| 35) Dibromofluoromethane    | 5.51   | 113 | 109740   | 51.69 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 103.38% |      |
| 50) Toluene-d8              | 8.72   | 98  | 383477   | 52.62 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 105.24% |      |
| 62) 4-Bromofluorobenzene    | 11.14  | 95  | 125340   | 45.78 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 91.56%  |      |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\  
Data File : VX005500.D  
Acq On : 23 Oct 2018 05:26  
Operator : JC/MD  
Sample : J5621-09  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 38 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
PA-SW-01

Quant Time: Oct 23 13:18:25 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X101218W.M  
Quant Title : SW846 8260  
QLast Update : Fri Oct 12 11:28:43 2018  
Response via : Initial Calibration

