

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005502.D
 Acq On : 23 Oct 2018 06:20
 Operator : JC/MD
 Sample : J5626-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PA-TWP-01

Quant Time: Oct 23 13:20:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	225049	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	324650	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	139868	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140101	56.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.18%	
35) Dibromofluoromethane	5.51	113	110913	51.58	ug/l	0.01
Spiked Amount 50.000			Recovery	=	103.16%	
50) Toluene-d8	8.72	98	380972	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
62) 4-Bromofluorobenzene	11.14	95	126014	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
Target Compounds						
16) Acetone	2.45	43	15679	10.857	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005502.D
Acq On : 23 Oct 2018 06:20
Operator : JC/MD
Sample : J5626-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PA-TWP-01

Quant Time: Oct 23 13:20:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

