

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005505.D
 Acq On : 23 Oct 2018 07:40
 Operator : JC/MD
 Sample : J5603-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RAV-GT106

Quant Time: Oct 23 13:21:24 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	221626	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	323858	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287623	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135349	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	142777	58.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	116.10%	
35) Dibromofluoromethane	5.51	113	111981	52.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.42%	
50) Toluene-d8	8.72	98	386385	52.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.94%	
62) 4-Bromofluorobenzene	11.14	95	124690	45.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005505.D
Acq On : 23 Oct 2018 07:40
Operator : JC/MD
Sample : J5603-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RAV-GT106

Quant Time: Oct 23 13:21:24 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

