

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102218\
 Data File : VX005513.D
 Acq On : 23 Oct 2018 11:13
 Operator : JC/MD
 Sample : J5538-18RE
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BLOIRE

Quant Time: Oct 23 13:24:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	218945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	337219	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	280425	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137766	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140331	57.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.50%	
35) Dibromofluoromethane	5.51	113	37148	16.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	33.26%	
50) Toluene-d8	8.72	98	376356	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
62) 4-Bromofluorobenzene	11.14	95	136141	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
Target Compounds						
16) Acetone	2.45	43	42550	30.285	ug/l	97
20) Methylene Chloride	2.85	84	17895	7.632	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102218\
Data File : VX005513.D
Acq On : 23 Oct 2018 11:13
Operator : JC/MD
Sample : J5538-18RE
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BL01RE

Quant Time: Oct 23 13:24:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

