


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102218\
Data File : VX005465.D
Acq On : 22 Oct 2018 12:01
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	10799	PASS
75	95	30	60	46.2	28856	PASS
95	95	100	100	100.0	62450	PASS
96	95	5	9	6.8	4263	PASS
173	174	0.00	2	1.4	833	PASS
174	95	50	100	94.5	59018	PASS
175	174	5	9	7.3	4306	PASS
176	174	95	101	97.8	57736	PASS
177	176	5	9	6.7	3855	PASS