

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102219\
 Data File : VX013172.D
 Acq On : 22 Oct 2019 12:55
 Operator : JC/SP
 Sample : K5460-02DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SEWER-OUTLET-02(DSN-001)

Manual Integrations
 APPROVED

MMDadoda
 10/23/2019 3:10:29 PM

Quant Time: Oct 23 07:51:36 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	25019	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	140679	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	125591	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.06	65	59342	30.13	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery	=	100.43%	
60) 4-Bromofluorobenzene	11.13	95	61775	28.25	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery	=	94.17%	
63) Toluene-d8	8.71	98	167909	30.06	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery	=	100.20%	

Target Compounds

					Ovalue
12) Methyl Acetate	2.76	43	10305	5.199 ug/l	99
15) Acetone	2.43	58	25460	93.232 ug/l	96
25) Chloroform	5.20	83	10146m	3.368 ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102219\
Data File : VX013172.D
Acq On : 22 Oct 2019 12:55
Operator : JC/SP
Sample : K5460-02DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SEWER-OUTLET-02(DSN-001)

Manual Integrations
APPROVED

MMDadoda
10/23/2019 3:10:29 PM

Quant Time: Oct 23 07:51:36 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 22 01:53:07 2019
Response via : Initial Calibration

