

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX102219\
 Data File : VX013168.D
 Acq On : 22 Oct 2019 09:44
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:47:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 M	Dichlorodifluoromethane	2.154	2.061	4.3	93	0.00
3 M	Chloromethane	1.878	1.753	6.7	94	0.00
4 M	Vinyl Chloride	1.941	1.811	6.7	94	0.00
5 M	Bromomethane	1.048	1.100	-5.0	91	0.00
6 M	Chloroethane	1.239	1.161	6.3	92	0.00
7 M	Trichlorofluoromethane	3.120	3.028	2.9	97	0.00
8 T	Diethyl Ether	1.065	0.993	6.8	94	0.00
9	1,1,2-Trichlorotrifluoroeth	1.906	1.850	2.9	100	0.00
10 M	1,1-Dichloroethene	1.840	1.709	7.1	93	0.00
11	Methyl Iodide	2.370	1.958	17.4	88	0.00
12	Methyl Acetate	2.377	2.186	8.0	90	0.00
13 M	Acrolein	0.354	0.250	29.4	66	0.00
14 M	Acrylonitrile	0.994	0.922	7.2	91	0.00
15 M	Acetone	0.327	0.362	-10.7	103	0.00
16 M	Carbon Disulfide	4.915	4.390	10.7	92	0.00
17	Allyl chloride	2.927	2.648	9.5	91	0.00
18 M	Methylene Chloride	2.131	1.980	7.1	94	0.00
19 M	trans-1,2-Dichloroethene	1.953	1.866	4.5	96	0.00
20 T	Diisopropyl ether	5.662	5.336	5.8	93	0.00
21 M	1,1-Dichloroethane	3.438	3.212	6.6	94	0.00
22 M	cis-1,2-Dichloroethene	2.227	2.136	4.1	95	0.00
23 M	tert-Butyl Alcohol	0.524	0.444	15.3	84	0.00
24 M	Methyl tert-Butyl Ether	6.347	5.968	6.0	95	0.00
25 M	Chloroform	3.613	3.482	3.6	96	0.00
26	Cyclohexane	3.016	2.867	4.9	97	0.00
27 s	1,2-Dichloroethane-d4	2.362	2.354	0.3	98	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
29	1,1-Dichloropropene	0.466	0.439	5.8	92	0.00
30 M	2-Butanone	0.244	0.251	-2.9	98	0.00
31	2,2-Dichloropropane	0.539	0.504	6.5	93	0.00
32 M	1,1,1-Trichloroethane	0.564	0.543	3.7	95	0.00
33 M	Carbon Tetrachloride	0.485	0.471	2.9	97	0.00
34 M	Benzene	1.342	1.291	3.8	93	0.00
35	Methacrylonitrile	0.251	0.240	4.4	95	0.00
36 M	1,2-Dichloroethane	0.507	0.491	3.2	94	0.00
37 M	Trichloroethene	0.372	0.371	0.3	98	0.00
38	Methylcyclohexane	0.568	0.557	1.9	99	0.00
39 M	1,2-Dichloropropane	0.329	0.313	4.9	93	0.00
40	Dibromomethane	0.235	0.231	1.7	96	0.00
41 M	Bromodichloromethane	0.474	0.457	3.6	95	0.00
42 M	Vinyl Acetate	0.848	0.804	5.2	92	0.00
43	Ethyl Acetate	0.446	0.420	5.8	94	0.00
44	Isopropyl Acetate	0.749	0.706	5.7	91	0.00
45 T	1,4-Dioxane	0.009	0.009#	0.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102219\
 Data File : VX013168.D
 Acq On : 22 Oct 2019 09:44
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:47:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.369	0.346	6.2	92	0.00
47	n-amyl Acetate	0.632	0.579	8.4	95	0.00
48 M	t-1,3-Dichloropropene	0.520	0.473	9.0	93	0.00
49 T	cis-1,3-Dichloropropene	0.566	0.528	6.7	93	0.00
50 M	1,1,2-Trichloroethane	0.347	0.342	1.4	96	0.00
51	Ethyl methacrylate	0.545	0.507	7.0	93	0.00
52	1,3-Dichloropropane	0.585	0.569	2.7	94	0.00
53 M	Dibromochloromethane	0.375	0.351	6.4	95	0.00
54 M	1,2-Dibromoethane	0.371	0.353	4.9	92	0.00
55 M	2-Chloroethyl vinyl ether	0.221	0.217	1.8	98	0.00
56 M	Bromoform	0.266	0.235	11.7	92	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
58 M	4-Methyl-2-Pentanone	0.506	0.479	5.3	91	0.00
59 M	2-Hexanone	0.396	0.380	4.0	92	0.00
60 S	4-Bromofluorobenzene	0.522	0.513	1.7	97	0.00
61 M	Tetrachloroethene	0.398	0.415	-4.3	98	0.00
62 M	Toluene	1.619	1.583	2.2	95	0.00
63 S	Toluene-d8	1.334	1.360	-1.9	98	0.00
64 M	Chlorobenzene	1.038	1.004	3.3	96	0.00
65	1,1,1,2-Tetrachloroethane	0.389	0.368	5.4	92	0.00
66 M	Ethyl Benzene	1.864	1.797	3.6	96	0.00
67 M	m/p-Xylenes	0.708	0.691	2.4	96	0.00
68 M	o-Xylene	0.676	0.639	5.5	96	0.00
69 M	Styrene	1.164	1.110	4.6	94	0.00
70	Isopropylbenzene	1.856	1.803	2.9	96	0.00
71 M	1,1,2,2-Tetrachloroethane	0.569	0.533	6.3	94	0.00
72	1,2,3-Trichloropropane	0.494	0.521	-5.5	94	0.00
73	Bromobenzene	0.445	0.433	2.7	99	0.00
74	n-propylbenzene	2.037	2.015	1.1	99	0.00
75	2-Chlorotoluene	1.274	1.226	3.8	95	0.00
76	1,3,5-Trimethylbenzene	1.558	1.464	6.0	95	0.00
77	t-1,4-Dichloro-2-butene	0.192	0.168	12.5	93	0.00
78	4-Chlorotoluene	1.456	1.391	4.5	97	0.00
79	tert-butylbenzene	1.500	1.406	6.3	98	0.00
80	1,2,4-Trimethylbenzene	1.556	1.506	3.2	99	0.00
81	sec-Butylbenzene	1.787	1.681	5.9	95	0.00
82	p-Isopropyltoluene	1.634	1.621	0.8	101	0.00
83 M	1,3-Dichlorobenzene	0.813	0.776	4.6	98	0.00
84 M	1,4-Dichlorobenzene	0.804	0.755	6.1	96	0.00
85	n-Butylbenzene	1.388	1.297	6.6	98	0.00
86 T	Hexachloroethane	0.285	0.257	9.8	94	0.00
87 M	1,2-Dichlorobenzene	0.789	0.754	4.4	95	0.00
88	1,2-Dibromo-3-Chloropropane	0.140	0.123	12.1	87	0.00
89	1,2,4-Trichlorobenzene	0.491	0.476	3.1	101	0.00
90	Hexachlorobutadiene	0.249	0.232	6.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102219\
Data File : VX013168.D
Acq On : 22 Oct 2019 09:44
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Oct 23 07:47:08 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 22 01:53:07 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.596	1.493	6.5	95	0.00
92	1,2,3-Trichlorobenzene	0.503	0.481	4.4	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0