

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX102219\
 Data File : VX013173.D
 Acq On : 22 Oct 2019 13:30
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:52:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	108	0.00
2 M	Dichlorodifluoromethane	2.154	1.997	7.3	100	0.00
3 M	Chloromethane	1.878	1.692	9.9	100	0.00
4 M	Vinyl Chloride	1.941	1.740	10.4	99	0.00
5 M	Bromomethane	1.048	1.144	-9.2	105	0.00
6 M	Chloroethane	1.239	1.062	14.3	93	0.00
7 M	Trichlorofluoromethane	3.120	2.893	7.3	102	0.00
8 T	Diethyl Ether	1.065	0.955	10.3	100	0.00
9	1,1,2-Trichlorotrifluoroeth	1.906	1.766	7.3	106	0.00
10 M	1,1-Dichloroethene	1.840	1.680	8.7	101	0.00
11	Methyl Iodide	2.370	2.001	15.6	99	0.00
12	Methyl Acetate	2.377	2.137	10.1	97	0.00
13 M	Acrolein	0.354	0.316	10.7	93	0.00
14 M	Acrylonitrile	0.994	0.875	12.0	95	0.00
15 M	Acetone	0.327	0.306	6.4	96	0.00
16 M	Carbon Disulfide	4.915	4.298	12.6	99	0.00
17	Allyl chloride	2.927	2.608	10.9	99	0.00
18 M	Methylene Chloride	2.131	1.870	12.2	98	0.00
19 M	trans-1,2-Dichloroethene	1.953	1.829	6.3	104	0.00
20 T	Diisopropyl ether	5.662	5.218	7.8	100	0.00
21 M	1,1-Dichloroethane	3.438	3.143	8.6	101	0.00
22 M	cis-1,2-Dichloroethene	2.227	2.006	9.9	99	0.00
23 M	tert-Butyl Alcohol	0.524	0.405	22.7	84	0.00
24 M	Methyl tert-Butyl Ether	6.347	5.894	7.1	104	0.00
25 M	Chloroform	3.613	3.321	8.1	101	0.00
26	Cyclohexane	3.016	2.745	9.0	102	0.00
27 s	1,2-Dichloroethane-d4	2.362	2.313	2.1	107	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	108	0.00
29	1,1-Dichloropropene	0.466	0.429	7.9	98	0.00
30 M	2-Butanone	0.244	0.223	8.6	96	0.00
31	2,2-Dichloropropane	0.539	0.515	4.5	104	0.00
32 M	1,1,1-Trichloroethane	0.564	0.532	5.7	103	0.00
33 M	Carbon Tetrachloride	0.485	0.454	6.4	103	0.00
34 M	Benzene	1.342	1.268	5.5	101	0.00
35	Methacrylonitrile	0.251	0.226	10.0	98	0.00
36 M	1,2-Dichloroethane	0.507	0.489	3.6	102	0.00
37 M	Trichloroethene	0.372	0.361	3.0	105	0.00
38	Methylcyclohexane	0.568	0.532	6.3	104	0.00
39 M	1,2-Dichloropropane	0.329	0.310	5.8	101	0.00
40	Dibromomethane	0.235	0.224	4.7	102	0.00
41 M	Bromodichloromethane	0.474	0.450	5.1	103	0.00
42 M	Vinyl Acetate	0.848	0.795	6.2	100	0.00
43	Ethyl Acetate	0.446	0.419	6.1	103	0.00
44	Isopropyl Acetate	0.749	0.695	7.2	99	0.00
45 T	1,4-Dioxane	0.009	0.008#	11.1	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102219\
 Data File : VX013173.D
 Acq On : 22 Oct 2019 13:30
 Operator : JC/SP
 Sample : VSTDCCC020
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC020

Quant Time: Oct 23 07:52:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 22 01:53:07 2019
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.369	0.343	7.0	100	0.00
47	n-amyl Acetate	0.632	0.577	8.7	104	0.00
48 M	t-1,3-Dichloropropene	0.520	0.482	7.3	104	0.00
49 T	cis-1,3-Dichloropropene	0.566	0.515	9.0	100	0.00
50 M	1,1,2-Trichloroethane	0.347	0.336	3.2	104	0.00
51	Ethyl methacrylate	0.545	0.520	4.6	105	0.00
52	1,3-Dichloropropane	0.585	0.564	3.6	103	0.00
53 M	Dibromochloromethane	0.375	0.340	9.3	101	0.00
54 M	1,2-Dibromoethane	0.371	0.347	6.5	100	0.00
55 M	2-Chloroethyl vinyl ether	0.221	0.230	-4.1	114	0.00
56 M	Bromoform	0.266	0.237	10.9	103	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
58 M	4-Methyl-2-Pentanone	0.506	0.461	8.9	97	0.00
59 M	2-Hexanone	0.396	0.351	11.4	95	0.00
60 S	4-Bromofluorobenzene	0.522	0.525	-0.6	111	0.00
61 M	Tetrachloroethene	0.398	0.389	2.3	102	0.00
62 M	Toluene	1.619	1.543	4.7	104	0.00
63 S	Toluene-d8	1.334	1.357	-1.7	109	0.00
64 M	Chlorobenzene	1.038	0.977	5.9	104	0.00
65	1,1,1,2-Tetrachloroethane	0.389	0.367	5.7	103	0.00
66 M	Ethyl Benzene	1.864	1.770	5.0	105	0.00
67 M	m/p-Xylenes	0.708	0.668	5.6	103	0.00
68 M	o-Xylene	0.676	0.640	5.3	107	0.00
69 M	Styrene	1.164	1.092	6.2	104	0.00
70	Isopropylbenzene	1.856	1.752	5.6	104	0.00
71 M	1,1,2,2-Tetrachloroethane	0.569	0.525	7.7	104	0.00
72	1,2,3-Trichloropropane	0.494	0.435	11.9	88	0.00
73	Bromobenzene	0.445	0.409	8.1	105	0.00
74	n-propylbenzene	2.037	1.994	2.1	109	0.00
75	2-Chlorotoluene	1.274	1.208	5.2	105	0.00
76	1,3,5-Trimethylbenzene	1.558	1.461	6.2	106	0.00
77	t-1,4-Dichloro-2-butene	0.192	0.162	15.6	100	0.00
78	4-Chlorotoluene	1.456	1.363	6.4	106	0.00
79	tert-butylbenzene	1.500	1.389	7.4	108	0.00
80	1,2,4-Trimethylbenzene	1.556	1.487	4.4	109	0.00
81	sec-Butylbenzene	1.787	1.649	7.7	104	0.00
82	p-Isopropyltoluene	1.634	1.577	3.5	110	0.00
83 M	1,3-Dichlorobenzene	0.813	0.762	6.3	108	0.00
84 M	1,4-Dichlorobenzene	0.804	0.765	4.9	109	0.00
85	n-Butylbenzene	1.388	1.285	7.4	108	0.00
86 T	Hexachloroethane	0.285	0.253	11.2	104	0.00
87 M	1,2-Dichlorobenzene	0.789	0.739	6.3	104	0.00
88	1,2-Dibromo-3-Chloropropane	0.140	0.115	17.9	90	0.00
89	1,2,4-Trichlorobenzene	0.491	0.485	1.2	115	0.00
90	Hexachlorobutadiene	0.249	0.241	3.2	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102219\
Data File : VX013173.D
Acq On : 22 Oct 2019 13:30
Operator : JC/SP
Sample : VSTDCCC020
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC020

Quant Time: Oct 23 07:52:49 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 22 01:53:07 2019
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	1.596	1.446	9.4	103	0.00
92	1,2,3-Trichlorobenzene	0.503	0.466	7.4	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0