


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102219\
Data File : VX013167.D
Acq On : 22 Oct 2019 09:12
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\624X102119W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Tue Oct 22 01:53:07 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1674

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.5	5356	PASS
75	95	30	60	49.2	15098	PASS
95	95	100	100	100.0	30693	PASS
96	95	5	9	6.9	2132	PASS
173	174	0.00	2	0.4	87	PASS
174	95	50	100	71.3	21897	PASS
175	174	5	9	7.8	1704	PASS
176	174	95	101	98.4	21541	PASS
177	176	5	9	6.4	1370	PASS