

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019021.D
 Acq On : 22 Oct 2020 13:00
 Operator : JC/SP
 Sample : L4214-08 5.0PPB L00
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LOQ-WATER-02-QT4-2020

Quant Time: Oct 23 10:50:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 23 09:59:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	239169	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	399647	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	366835	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	178076	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	140937	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
35) Dibromofluoromethane	5.46	113	119384	47.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.36%	
50) Toluene-d8	8.70	98	461943	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
62) 4-Bromofluorobenzene	11.12	95	157693	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
Target Compounds						
5) Bromomethane	1.64	94	16276	5.705	ug/l	95
10) Methyl Iodide	2.50	142	10853	3.847	ug/l	98
11) Tert butyl alcohol	3.04	59	15224	26.238	ug/l #	89
13) Acrolein	2.28	56	6626	28.823	ug/l	100
31) Cyclohexane	5.55	56	16496	4.917	ug/l	92
81) trans-1,4-Dichloro-2-buten	11.06	75	6171	4.417	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019021.D
Acq On : 22 Oct 2020 13:00
Operator : JC/SP
Sample : L4214-08 5.0PPB L00
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LOQ-WATER-02-QT4-2020

Quant Time: Oct 23 10:50:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 23 09:59:01 2020
Response via : Initial Calibration

