

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019032.D
 Acq On : 22 Oct 2020 17:07
 Operator : JC/SP
 Sample : L4417-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16

Quant Time: Oct 23 09:24:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	266742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	464933	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	456003	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231116	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	166386	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
35) Dibromofluoromethane	5.46	113	145113	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	8.70	98	585131	52.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.18%	
62) 4-Bromofluorobenzene	11.12	95	211958	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
Target Compounds						
16) Acetone	2.42	43	2871	2.638	ug/l	89
19) Methyl tert-butyl Ether	3.16	73	8097	0.985	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019032.D
Acq On : 22 Oct 2020 17:07
Operator : JC/SP
Sample : L4417-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16

Quant Time: Oct 23 09:24:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

