

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019033.D
 Acq On : 22 Oct 2020 17:30
 Operator : JC/SP
 Sample : L4417-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14

Quant Time: Oct 23 09:27:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	263230	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	456666	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	446405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	240179	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	163639	51.90	ug/l	0.00
Spiked Amount			Recovery	=	103.80%	
35) Dibromofluoromethane	5.46	113	143483	49.63	ug/l	0.00
Spiked Amount			Recovery	=	99.26%	
50) Toluene-d8	8.70	98	576227	52.23	ug/l	0.00
Spiked Amount			Recovery	=	104.46%	
62) 4-Bromofluorobenzene	11.12	95	217658	55.46	ug/l	0.00
Spiked Amount			Recovery	=	110.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.42	43	4942	4.602	ug/l	95
31) Cyclohexane	5.54	56	18509	5.013	ug/l	93
39) Methylcyclohexane	7.44	83	15027	3.273	ug/l	93
67) Ethyl Benzene	10.23	91	207273	12.842	ug/l	100
68) m/p-Xylenes	10.35	106	11065	1.790	ug/l	100
73) Isopropylbenzene	11.00	105	1696453	98.186	ug/l	99
78) n-propylbenzene	11.34	91	2485094	126.932	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	168426	11.805	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	125327	8.772	ug/l	99
85) sec-Butylbenzene	11.93	105	170541	10.027	ug/l	88
89) n-Butylbenzene	12.37	91	190911	13.420	ug/l #	70
95) Naphthalene	13.82	128	274578	15.983	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019033.D
Acq On : 22 Oct 2020 17:30
Operator : JC/SP
Sample : L4417-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14

Quant Time: Oct 23 09:27:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

