

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019038.D
 Acq On : 22 Oct 2020 19:22
 Operator : JC/SP
 Sample : L4417-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6D

Quant Time: Oct 23 09:43:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	263011	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	457257	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439499	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226090	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	162063	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	5.46	113	142486	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	8.70	98	566780	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.12	95	208534	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.01	59	3958	6.203	ug/l #	94
16) Acetone	2.42	43	3795	3.537	ug/l	94
17) Carbon Disulfide	2.55	76	6358	0.861	ug/l	96
19) Methyl tert-butyl Ether	3.17	73	10845	1.337	ug/l	100
20) Methylene Chloride	2.83	84	820	0.274	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019038.D
Acq On : 22 Oct 2020 19:22
Operator : JC/SP
Sample : L4417-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6D

Quant Time: Oct 23 09:43:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

