

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019051.D
 Acq On : 23 Oct 2020 01:20
 Operator : JC/SP
 Sample : L4417-22
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Oct 23 08:56:55 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	264224	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	461940	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	439471	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	219638	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	163793	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	5.46	113	143366	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
50) Toluene-d8	8.70	98	570980	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.12	95	202966	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
Target Compounds						
16) Acetone	2.43	43	2529	2.346	ug/l	93
20) Methylene Chloride	2.84	84	1019	0.340	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019051.D
Acq On : 23 Oct 2020 01:20
Operator : JC/SP
Sample : L4417-22
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Oct 23 08:56:55 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

