

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019057.D
 Acq On : 23 Oct 2020 03:34
 Operator : JC/SP
 Sample : L4417-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1D

Manual Integrations
 APPROVED

MMDadoda
 10/23/2020 1:20:26 PM

Quant Time: Oct 23 09:27:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	318340	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	548097	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	522235	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	268106	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	185512	48.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.30%	
35) Dibromofluoromethane	5.46	113	169368	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.70	98	667877	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	11.12	95	241394	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.00	59	752468	974.325	ug/l	100
16) Acetone	2.42	43	6001	4.620	ug/l	99
17) Carbon Disulfide	2.56	76	30878	3.455	ug/l	99
19) Methyl tert-butyl Ether	3.17	73	5933285	604.518	ug/l	100
31) Cyclohexane	5.54	56	1354m	0.303	ug/l	
40) Benzene	6.11	78	13916	0.961	ug/l	99
52) Toluene	8.76	92	63404	6.628	ug/l	99
67) Ethyl Benzene	10.24	91	43934	2.327	ug/l	99
68) m/p-Xylenes	10.34	106	25449	3.518	ug/l	100
69) o-Xylene	10.68	106	7247	1.047	ug/l	100
73) Isopropylbenzene	11.00	105	9339	0.484	ug/l	99
78) n-propylbenzene	11.34	91	17365	0.795	ug/l	98
80) 1,3,5-Trimethylbenzene	11.49	105	9349	0.587	ug/l	98
84) 1,2,4-Trimethylbenzene	11.79	105	32516	2.039	ug/l	97
95) Naphthalene	13.82	128	36310	1.893	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019057.D
Acq On : 23 Oct 2020 03:34
Operator : JC/SP
Sample : L4417-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1D

Manual Integrations
APPROVED

MMDadoda
10/23/2020 1:20:26 PM

Quant Time: Oct 23 09:27:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

