

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102220\
 Data File : VX019059.D
 Acq On : 23 Oct 2020 04:19
 Operator : JC/SP
 Sample : PB132537TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132537TB

Quant Time: Oct 23 09:47:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	310539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	525427	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	513070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	259721	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	181485	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	
35) Dibromofluoromethane	5.46	113	163834	49.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.50%	
50) Toluene-d8	8.70	98	648102	51.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.10%	
62) 4-Bromofluorobenzene	11.12	95	236884	52.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.92%	

Target Compounds

					Qvalue
16) Acetone	2.42	43	9773	7.714	ug/l 98
18) Methyl Acetate	2.75	43	2929	0.932	ug/l 93
20) Methylene Chloride	2.83	84	5424	1.538	ug/l 95
43) Isopropyl Acetate	6.41	43	63801	9.362	ug/l 98
52) Toluene	8.77	92	17180	1.873	ug/l 98
67) Ethyl Benzene	10.24	91	35306	1.903	ug/l 98
68) m/p-Xylenes	10.34	106	9420	1.326	ug/l 98
69) o-Xylene	10.68	106	1554	0.228	ug/l 79
73) Isopropylbenzene	11.00	105	8551	0.458	ug/l 98
95) Naphthalene	13.82	128	33750	1.817	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102220\
Data File : VX019059.D
Acq On : 23 Oct 2020 04:19
Operator : JC/SP
Sample : PB132537TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
PB132537TB

Quant Time: Oct 23 09:47:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

