


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102220\
 Data File : VX019013.D
 Acq On : 22 Oct 2020 09:27
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Title : SW846 8260
 Last Update : Thu Oct 22 17:51:42 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.6	9278	PASS
75	95	30	60	48.5	27173	PASS
95	95	100	100	100.0	55989	PASS
96	95	5	9	6.7	3753	PASS
173	174	0.00	2	1.1	537	PASS
174	95	50	100	86.8	48626	PASS
175	174	5	9	7.0	3385	PASS
176	174	95	101	95.3	46344	PASS
177	176	5	9	6.4	2979	PASS