

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
 Data File : VX024870.D
 Acq On : 22 Oct 2021 14:38
 Operator : JC/MD
 Sample : M4291-01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P001-TW001-05

Quant Time: Oct 23 03:54:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 23 03:44:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	67720	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	60573	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	29308	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	11674	5.872	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	117.400%
7) Chloroethane-d5	1.672	69	10176	5.681	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.600%
11) 1,1-Dichloroethene-d2	2.306	63	32189	4.343	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	86.800%
20) 2-Butanone-d5	4.477	46	38802	53.382	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.760%
24) Chloroform-d	5.068	84	54612	5.706	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	114.200%
26) 1,2-Dichloroethane-d4	5.964	65	28497	5.448	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.000%
32) Benzene-d6	5.983	84	87140	6.174	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	123.400%
36) 1,2-Dichloropropane-d6	7.312	67	25605	5.686	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.800%
41) Toluene-d8	8.653	98	87310	6.170	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	123.400%
43) trans-1,3-Dichloroprop...	8.952	79	9994	4.836	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.800%
46) 2-Hexanone-d5	9.384	63	41159	55.335	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.660%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	19585	5.691	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.800%
66) 1,2-Dichlorobenzene-d4	12.323	152	29682	6.010	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.200%#
Target Compounds						
42) Toluene	8.726	91	2260	0.109	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
Data File : VX024870.D
Acq On : 22 Oct 2021 14:38
Operator : JC/MD
Sample : M4291-01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P001-TW001-05

Quant Time: Oct 23 03:54:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 23 03:44:55 2021
Response via : Initial Calibration

